

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
 Data File : VY022639.D
 Acq On : 10 Jun 2025 10:20
 Operator : SY/MD
 Sample : VY0610SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0610SBL01

Quant Time: Jun 11 01:46:51 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	238087	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	437151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	353511	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	122259	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	137102	51.487	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.980%
35) Dibromofluoromethane	7.640	113	133210	51.053	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.100%
50) Toluene-d8	10.109	98	529793	50.250	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.408	95	127448	40.572	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
Data File : VY022639.D
Acq On : 10 Jun 2025 10:20
Operator : SY/MD
Sample : VY0610SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0610SBL01

Quant Time: Jun 11 01:46:51 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

