

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
 Data File : VY022643.D
 Acq On : 10 Jun 2025 12:22
 Operator : SY/MD
 Sample : Q2274-02
 Misc : 4.89g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-13-MHP-VOC

Quant Time: Jun 11 01:49:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	257370	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	479196	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	386407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	134064	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	145006	50.375	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.740%
35) Dibromofluoromethane	7.640	113	146332	51.161	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.320%
50) Toluene-d8	10.109	98	578672	50.071	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.140%
62) 4-Bromofluorobenzene	12.408	95	142297	41.324	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
Data File : VY022643.D
Acq On : 10 Jun 2025 12:22
Operator : SY/MD
Sample : Q2274-02
Misc : 4.89g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-13-MHP-VOC

Quant Time: Jun 11 01:49:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

