

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
 Data File : VY022645.D
 Acq On : 10 Jun 2025 13:09
 Operator : SY/MD
 Sample : Q2273-07
 Misc : 6.41g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-6-VOC

Quant Time: Jun 11 01:50:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	251717	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	460760	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	370629	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	130026	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	137580	48.868	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.740%
35) Dibromofluoromethane	7.640	113	137673	50.060	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.120%
50) Toluene-d8	10.109	98	563398	50.700	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.400%
62) 4-Bromofluorobenzene	12.402	95	131016	39.570	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
Data File : VY022645.D
Acq On : 10 Jun 2025 13:09
Operator : SY/MD
Sample : Q2273-07
Misc : 6.41g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-6-VOC

Quant Time: Jun 11 01:50:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

