

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
 Data File : VY022647.D
 Acq On : 10 Jun 2025 13:56
 Operator : SY/MD
 Sample : Q2259-01
 Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-36-060525

Quant Time: Jun 11 01:51:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	226891	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	430118	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	337782	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	119411	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	124313	48.987	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.980%
35) Dibromofluoromethane	7.634	113	125182	48.760	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.520%
50) Toluene-d8	10.109	98	511879	49.345	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.402	95	121940	39.453	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061025\
Data File : VY022647.D
Acq On : 10 Jun 2025 13:56
Operator : SY/MD
Sample : Q2259-01
Misc : 4.50g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-36-060525

Quant Time: Jun 11 01:51:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

