

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061124\
 Data File : VY018563.D
 Acq On : 11 Jun 2024 17:36
 Operator : SY/MD
 Sample : P2812-01
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-05-061024

Quant Time: Jun 12 03:04:32 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 04 07:11:56 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	520	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	893	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	481	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	21	50.000	ug/l	# 0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	906	88.719	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	177.440%#
35) Dibromofluoromethane	7.710	113	577	50.954	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.900%
50) Toluene-d8	10.185	98	1411	31.321	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	62.640%
62) 4-Bromofluorobenzene	12.483	95	67	4.777	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	9.560%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061124\
Data File : VY018563.D
Acq On : 11 Jun 2024 17:36
Operator : SY/MD
Sample : P2812-01
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-05-061024

Quant Time: Jun 12 03:04:32 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060324S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 04 07:11:56 2024
Response via : Initial Calibration

