

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061225\
 Data File : VY022678.D
 Acq On : 12 Jun 2025 16:55
 Operator : SY/MD
 Sample : Q2295-06
 Misc : 6.43g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP02-MH5-VOC

Quant Time: Jun 13 05:43:28 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	236331	50.000	ug/l	-0.01
34) 1,4-Difluorobenzene	8.610	114	433690	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.408	117	359868	50.000	ug/l	-0.01
72) 1,4-Dichlorobenzene-d4	13.340	152	135204	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	129727	49.079	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.160%
35) Dibromofluoromethane	7.628	113	130787	50.524	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.040%
50) Toluene-d8	10.103	98	534269	51.079	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.160%
62) 4-Bromofluorobenzene	12.402	95	134115	43.035	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061225\
Data File : VY022678.D
Acq On : 12 Jun 2025 16:55
Operator : SY/MD
Sample : Q2295-06
Misc : 6.43g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP02-MH5-VOC

Quant Time: Jun 13 05:43:28 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

