


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061225\
Data File : VY022665.D
Acq On : 12 Jun 2025 08:56
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Title : SW846 8260
Last Update : Tue Jun 10 15:26:14 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	23.5	18174	PASS
75	95	30	60	57.0	44051	PASS
95	95	100	100	100.0	77347	PASS
96	95	5	9	6.5	5015	PASS
173	174	0.00	2	1.4	950	PASS
174	95	50	100	87.2	67413	PASS
175	174	5	9	8.0	5374	PASS
176	174	95	101	98.2	66208	PASS
177	176	5	9	6.6	4399	PASS