

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061325\
 Data File : VY022683.D
 Acq On : 13 Jun 2025 09:13
 Operator : SY/MD
 Sample : VY0613SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0613SBL01

Quant Time: Jun 14 02:00:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	169791	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	313555	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	249938	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	87111	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	92485	48.701	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.400%
35) Dibromofluoromethane	7.634	113	95464	51.008	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.020%
50) Toluene-d8	10.103	98	377974	49.982	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.960%
62) 4-Bromofluorobenzene	12.402	95	87831	38.981	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061325\
Data File : VY022683.D
Acq On : 13 Jun 2025 09:13
Operator : SY/MD
Sample : VY0613SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0613SBL01

Quant Time: Jun 14 02:00:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

