

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061325\
 Data File : VY022689.D
 Acq On : 13 Jun 2025 11:55
 Operator : SY/MD
 Sample : Q2311-06
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP04-MH2MH3-VOC

Quant Time: Jun 14 02:03:30 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	178619	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	333609	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	267351	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	95264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	100535	50.324	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	100.640%	
35) Dibromofluoromethane	7.634	113	99504	49.971	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	99.940%	
50) Toluene-d8	10.103	98	399261	49.623	ug/l	0.00
Spiked Amount	50.000	Range 58 - 134	Recovery	=	99.240%	
62) 4-Bromofluorobenzene	12.401	95	98682	41.164	ug/l	0.00
Spiked Amount	50.000	Range 30 - 143	Recovery	=	82.320%	
Target Compounds						
						Qvalue
20) Methylene Chloride	4.610	84	8640	3.749	ug/l	# 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061325\
Data File : VY022689.D
Acq On : 13 Jun 2025 11:55
Operator : SY/MD
Sample : Q2311-06
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP04-MH2MH3-VOC

Quant Time: Jun 14 02:03:30 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

