

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061625\
 Data File : VY022706.D
 Acq On : 16 Jun 2025 14:59
 Operator : SY/MD
 Sample : Q2305-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-06122025RE

Quant Time: Jun 17 05:54:57 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	10768	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	16222	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	8470	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	1790	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	3350	27.816	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	55.640%
35) Dibromofluoromethane	7.634	113	3133	32.357	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	64.720%
50) Toluene-d8	10.109	98	15982	40.850	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	81.700%
62) 4-Bromofluorobenzene	12.408	95	2214	18.993	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	37.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061625\
Data File : VY022706.D
Acq On : 16 Jun 2025 14:59
Operator : SY/MD
Sample : Q2305-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-06122025RE

Quant Time: Jun 17 05:54:57 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

