

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061724\
 Data File : VY018611.D
 Acq On : 17 Jun 2024 15:44
 Operator : SY/MD
 Sample : P2889-03RE
 Misc : 4.77g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WCS-TPJ-VOCRE

Quant Time: Jun 18 01:25:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	49655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	93718	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	83175	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	33301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	28832	57.673	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.340%
35) Dibromofluoromethane	7.704	113	28895	50.543	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.080%
50) Toluene-d8	10.173	98	111037	50.505	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	101.000%
62) 4-Bromofluorobenzene	12.477	95	35545	51.200	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.400%
Target Compounds						
20) Methylene Chloride	4.692	84	9970	8.462	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061724\
Data File : VY018611.D
Acq On : 17 Jun 2024 15:44
Operator : SY/MD
Sample : P2889-03RE
Misc : 4.77g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WCS-TPJ-VOCRE

Quant Time: Jun 18 01:25:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

