



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061724\
Data File : VY018604.D
Acq On : 17 Jun 2024 09:18
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Title : SW846 8260
Last Update : Fri Jun 14 07:08:24 2024



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.1	18981	PASS
75	95	30	60	56.7	51067	PASS
95	95	100	100	100.0	90019	PASS
96	95	5	9	6.7	5992	PASS
173	174	0.00	2	1.0	743	PASS
174	95	50	100	79.9	71907	PASS
175	174	5	9	7.5	5404	PASS
176	174	95	101	97.3	69957	PASS
177	176	5	9	6.6	4602	PASS