

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061725\
 Data File : VY022713.D
 Acq On : 17 Jun 2025 09:20
 Operator : SY/MD
 Sample : VY0617SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY0617SBL01

Quant Time: Jun 18 04:37:05 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	105503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	194865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	159398	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	57457	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	61284	51.936	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.880%
35) Dibromofluoromethane	7.640	113	60464	51.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.960%
50) Toluene-d8	10.109	98	236579	50.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.680%
62) 4-Bromofluorobenzene	12.408	95	57392	40.986	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061725\
Data File : VY022713.D
Acq On : 17 Jun 2025 09:20
Operator : SY/MD
Sample : VY0617SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY0617SBL01

Quant Time: Jun 18 04:37:05 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

