

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061725\
 Data File : VY022719.D
 Acq On : 17 Jun 2025 13:14
 Operator : SY/MD
 Sample : Q2336-03
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WBR-2

Quant Time: Jun 18 04:40:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	99565	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	182566	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	151714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	54317	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	58484	52.519	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 105.040%	
35) Dibromofluoromethane	7.640	113	57209	52.500	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 105.000%	
50) Toluene-d8	10.109	98	226051	51.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 102.680%	
62) 4-Bromofluorobenzene	12.407	95	53382	40.691	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 81.380%	
Target Compounds						
20) Methylene Chloride	4.616	84	5103	3.972	ug/l	Qvalue 91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061725\
Data File : VY022719.D
Acq On : 17 Jun 2025 13:14
Operator : SY/MD
Sample : Q2336-03
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WBR-2

Quant Time: Jun 18 04:40:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

