

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061825\
 Data File : VY022732.D
 Acq On : 18 Jun 2025 11:52
 Operator : SY/MD
 Sample : Q2338-01RE
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-03-061625RE

Quant Time: Jun 19 03:59:07 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	17034	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	27177	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	17922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	4868	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	7781	40.842	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	81.680%
35) Dibromofluoromethane	7.634	113	8472	52.227	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.460%
50) Toluene-d8	10.109	98	30906	47.152	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.300%
62) 4-Bromofluorobenzene	12.408	95	5873	30.073	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	60.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061825\
Data File : VY022732.D
Acq On : 18 Jun 2025 11:52
Operator : SY/MD
Sample : Q2338-01RE
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-03-061625RE

Quant Time: Jun 19 03:59:07 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

