

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061825\
 Data File : VY022734.D
 Acq On : 18 Jun 2025 12:39
 Operator : SY/MD
 Sample : Q2336-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WBR-1

Quant Time: Jun 19 03:59:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	86256	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	157947	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	131320	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	47143	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	48531	50.306	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.620%
35) Dibromofluoromethane	7.634	113	50372	53.431	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.860%
50) Toluene-d8	10.103	98	191320	50.224	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.440%
62) 4-Bromofluorobenzene	12.401	95	47046	41.451	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061825\
Data File : VY022734.D
Acq On : 18 Jun 2025 12:39
Operator : SY/MD
Sample : Q2336-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WBR-1

Quant Time: Jun 19 03:59:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

