



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061825\
Data File : VY022727.D
Acq On : 18 Jun 2025 09:15
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Title : SW846 8260
Last Update : Tue Jun 10 15:26:14 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.8	16417	PASS
75	95	30	60	59.2	39107	PASS
95	95	100	100	100.0	66101	PASS
96	95	5	9	6.9	4565	PASS
173	174	0.00	2	1.5	852	PASS
174	95	50	100	84.4	55765	PASS
175	174	5	9	7.6	4258	PASS
176	174	95	101	96.5	53821	PASS
177	176	5	9	6.2	3331	PASS