

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061925\
 Data File : VY022747.D
 Acq On : 19 Jun 2025 11:51
 Operator : SY/MD
 Sample : Q2357-01RE
 Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-06182025RE

Quant Time: Jun 20 02:57:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.701	168	4987	50.000	ug/l	#-0.01
34) 1,4-Difluorobenzene	8.616	114	7080	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	4641	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	684	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	2316	41.523	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.040%
35) Dibromofluoromethane	7.628	113	1118	26.456	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	52.920%#
50) Toluene-d8	10.103	98	7339	42.980	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	85.960%
62) 4-Bromofluorobenzene	12.408	95	950	18.673	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	37.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061925\
Data File : VY022747.D
Acq On : 19 Jun 2025 11:51
Operator : SY/MD
Sample : Q2357-01RE
Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-06182025RE

Quant Time: Jun 20 02:57:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

