

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061925\
 Data File : VY022749.D
 Acq On : 19 Jun 2025 12:38
 Operator : SY/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_Y/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VIBLK

Quant Time: Jun 20 02:57:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	782	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.610	114	1128	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	388	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	133	50.000	ug/l	# 0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	494	56.481	ug/l	0.00
Spiked Amount	50.000		Recovery	=	112.960%	
35) Dibromofluoromethane	7.622	113	330	49.014	ug/l	-0.02
Spiked Amount	50.000		Recovery	=	98.020%	
50) Toluene-d8	10.109	98	888	32.641	ug/l	0.00
Spiked Amount	50.000		Recovery	=	65.280%	
62) 4-Bromofluorobenzene	12.402	95	59	7.279	ug/l	0.00
Spiked Amount	50.000		Recovery	=	14.560%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061925\
Data File : VY022749.D
Acq On : 19 Jun 2025 12:38
Operator : SY/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_Y/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VIBLK

Quant Time: Jun 20 02:57:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

