


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY061925\
Data File : VY022741.D
Acq On : 19 Jun 2025 08:38
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Title : SW846 8260
Last Update : Tue Jun 10 15:26:14 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.3	16847	PASS
75	95	30	60	58.6	39093	PASS
95	95	100	100	100.0	66677	PASS
96	95	5	9	6.2	4105	PASS
173	174	0.00	2	1.4	794	PASS
174	95	50	100	87.2	58125	PASS
175	174	5	9	8.4	4873	PASS
176	174	95	101	96.7	56203	PASS
177	176	5	9	6.5	3681	PASS