



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062024\
Data File : VY018647.D
Acq On : 20 Jun 2024 09:25
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Title : SW846 8260
Last Update : Fri Jun 14 07:08:24 2024



AutoFind: Scans 1747, 1748, 1749; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.4	8805	PASS
75	95	30	60	54.6	23592	PASS
95	95	100	100	100.0	43245	PASS
96	95	5	9	6.5	2825	PASS
173	174	0.00	2	1.1	391	PASS
174	95	50	100	83.5	36096	PASS
175	174	5	9	8.2	2958	PASS
176	174	95	101	96.8	34955	PASS
177	176	5	9	6.8	2378	PASS