

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062025\
 Data File : VY022767.D
 Acq On : 20 Jun 2025 16:42
 Operator : SY/MD
 Sample : Q2380-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VNJ-211

Quant Time: Jun 21 02:42:00 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 10 15:26:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	99123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	183950	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	159048	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	59131	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	63013	56.838	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	113.680%
35) Dibromofluoromethane	7.634	113	59304	54.013	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.020%
50) Toluene-d8	10.109	98	228063	51.407	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.820%
62) 4-Bromofluorobenzene	12.402	95	58759	44.452	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	88.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062025\
Data File : VY022767.D
Acq On : 20 Jun 2025 16:42
Operator : SY/MD
Sample : Q2380-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VNJ-211

Quant Time: Jun 21 02:42:00 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 10 15:26:14 2025
Response via : Initial Calibration

