



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062025\
Data File : VY022758.D
Acq On : 20 Jun 2025 11:13
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060225S.M
Title : SW846 8260
Last Update : Tue Jun 10 15:26:14 2025



AutoFind: Scans 1752, 1753, 1754; Background Corrected with Scan 1744

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	25.5	15272	PASS
75	95	30	60	59.3	35440	PASS
95	95	100	100	100.0	59811	PASS
96	95	5	9	6.2	3694	PASS
173	174	0.00	2	1.7	862	PASS
174	95	50	100	83.7	50048	PASS
175	174	5	9	7.8	3922	PASS
176	174	95	101	97.4	48760	PASS
177	176	5	9	6.5	3185	PASS