

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014338.D
 Acq On : 21 Jun 2023 16:08
 Operator : KP/MD
 Sample : 03337-16
 Misc : 4.32g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-15

Quant Time: Jun 22 01:57:37 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	185549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	293902	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	278932	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	133301	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	124944	48.357	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	96.720%
35) Dibromofluoromethane	7.715	113	96910	46.496	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.000%
50) Toluene-d8	10.178	98	354508	45.495	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.980%
62) 4-Bromofluorobenzene	12.483	95	115454	41.709	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014338.D
Acq On : 21 Jun 2023 16:08
Operator : KP/MD
Sample : 03337-16
Misc : 4.32g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-15

Quant Time: Jun 22 01:57:37 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

