

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014345.D
 Acq On : 21 Jun 2023 18:52
 Operator : KP/MD
 Sample : 03358-02
 Misc : 5.14g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1-(8-10)

Quant Time: Jun 22 01:58:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	183052	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	290279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	279096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	137693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	130453	51.178	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.360%	
35) Dibromofluoromethane	7.716	113	95544	46.412	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 92.820%	
50) Toluene-d8	10.179	98	350453	45.536	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 91.080%	
62) 4-Bromofluorobenzene	12.483	95	117394	42.940	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 85.880%	
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	30002	9.313	ug/l	93
25) 2-Butanone	6.984	43	16022	16.940	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014345.D
Acq On : 21 Jun 2023 18:52
Operator : KP/MD
Sample : 03358-02
Misc : 5.14g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1-(8-10)

Quant Time: Jun 22 01:58:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

