

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014346.D
 Acq On : 21 Jun 2023 19:15
 Operator : KP/MD
 Sample : 03358-03
 Misc : 3.62g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-1-(34-36)

Quant Time: Jun 22 01:58:55 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	192463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	307502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	296562	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	145019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	133333	49.750	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.500%
35) Dibromofluoromethane	7.722	113	101186	46.400	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.800%
50) Toluene-d8	10.179	98	367722	45.103	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.200%
62) 4-Bromofluorobenzene	12.483	95	124235	42.897	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	85.800%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	24567	6.399	ug/l	83
25) 2-Butanone	6.990	43	7772	7.815	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014346.D
Acq On : 21 Jun 2023 19:15
Operator : KP/MD
Sample : 03358-03
Misc : 3.62g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-1-(34-36)

Quant Time: Jun 22 01:58:55 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

