

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014347.D
 Acq On : 21 Jun 2023 19:38
 Operator : KP/MD
 Sample : 03358-04
 Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-4-(0-2)

Quant Time: Jun 22 01:59:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	185156	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	291650	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	275709	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	133838	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	126350	49.005	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.000%
35) Dibromofluoromethane	7.722	113	95461	46.154	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.300%
50) Toluene-d8	10.185	98	350935	45.384	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.760%
62) 4-Bromofluorobenzene	12.483	95	123589	44.993	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	89.980%
Target Compounds						
					Qvalue	
16) Acetone	3.954	43	9021	11.607	ug/l #	88
20) Methylene Chloride	4.722	84	27150	7.925	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014347.D
Acq On : 21 Jun 2023 19:38
Operator : KP/MD
Sample : 03358-04
Misc : 5.23g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-4-(0-2)

Quant Time: Jun 22 01:59:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

