

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014358.D
 Acq On : 22 Jun 2023 01:04
 Operator : KP/MD
 Sample : 03358-09
 Misc : 3.91g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-6-(24-28)

Quant Time: Jun 22 02:01:18 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	188395	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	295876	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	282283	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	137965	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	125904	47.992	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.980%
35) Dibromofluoromethane	7.728	113	95836	45.673	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.340%
50) Toluene-d8	10.185	98	356000	45.381	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.760%
62) 4-Bromofluorobenzene	12.489	95	115730	41.530	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.060%
Target Compounds						Qvalue
20) Methylene Chloride	4.729	84	29853	8.876	ug/l	89
25) 2-Butanone	6.996	43	6881	7.069	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014358.D
Acq On : 22 Jun 2023 01:04
Operator : KP/MD
Sample : 03358-09
Misc : 3.91g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-6-(24-28)

Quant Time: Jun 22 02:01:18 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

