

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014359.D
 Acq On : 22 Jun 2023 01:27
 Operator : KP/MD
 Sample : 03358-12
 Misc : 5.33g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 DUP-01

Quant Time: Jun 22 04:51:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	181882	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	286849	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	277440	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	133657	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	125116	49.400	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.800%
35) Dibromofluoromethane	7.728	113	93320	45.874	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.740%
50) Toluene-d8	10.185	98	346484	45.558	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.120%
62) 4-Bromofluorobenzene	12.489	95	116845	43.250	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.500%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.728	84	22916	6.266	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014359.D
Acq On : 22 Jun 2023 01:27
Operator : KP/MD
Sample : 03358-12
Misc : 5.33g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
DUP-01

Quant Time: Jun 22 04:51:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

