

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014360.D
 Acq On : 22 Jun 2023 01:51
 Operator : KP/MD
 Sample : 03330-05
 Misc : 4.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-MC03-GRAB-6

Quant Time: Jun 22 04:51:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	180967	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.697	114	287756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	261249	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	111117	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	122511	48.616	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.240%
35) Dibromofluoromethane	7.728	113	94145	46.134	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	92.260%
50) Toluene-d8	10.185	98	344352	45.135	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.280%
62) 4-Bromofluorobenzene	12.489	95	103012	38.009	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.020%
Target Compounds						
29) Tetrahydrofuran	7.368	42	3773	6.363	ug/l	Qvalue # 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014360.D
Acq On : 22 Jun 2023 01:51
Operator : KP/MD
Sample : 03330-05
Misc : 4.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-MC03-GRAB-6

Quant Time: Jun 22 04:51:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

