

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
 Data File : VY014367.D
 Acq On : 22 Jun 2023 07:18
 Operator : KP/MD
 Sample : 03362-06
 Misc : 14.14g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 46 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TOTAL-VOC-SB-32-37-SB-32-GRAB-5

Quant Time: Jun 22 07:31:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 09 18:07:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	178374	50.000	ug/l	0.01
34) 1,4-Difluorobenzene	8.703	114	282687	50.000	ug/l	0.01
63) Chlorobenzene-d5	11.502	117	265841	50.000	ug/l	0.01
72) 1,4-Dichlorobenzene-d4	13.434	152	112952	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	129614	52.182	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	104.360%
35) Dibromofluoromethane	7.734	113	96903	48.337	ug/l	0.02
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.680%
50) Toluene-d8	10.191	98	343694	45.857	ug/l	0.01
Spiked Amount	50.000	Range	58 - 134	Recovery	=	91.720%
62) 4-Bromofluorobenzene	12.489	95	103021	38.694	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062123\
Data File : VY014367.D
Acq On : 22 Jun 2023 07:18
Operator : KP/MD
Sample : 03362-06
Misc : 14.14g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 46 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TOTAL-VOC-SB-32-37-SB-32-GRAB-5

Quant Time: Jun 22 07:31:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y060923S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 09 18:07:13 2023
Response via : Initial Calibration

