

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062425\
 Data File : VY022794.D
 Acq On : 24 Jun 2025 13:55
 Operator : SY/MD
 Sample : Q2392-02
 Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S-1A

Quant Time: Jun 25 01:25:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	337439	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	618561	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	592770	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	269981	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	193901	51.485	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.980%
35) Dibromofluoromethane	7.628	113	191589	50.930	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.860%
50) Toluene-d8	10.103	98	736519	49.340	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.680%
62) 4-Bromofluorobenzene	12.401	95	270040	56.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	112.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062425\
Data File : VY022794.D
Acq On : 24 Jun 2025 13:55
Operator : SY/MD
Sample : Q2392-02
Misc : 5.16g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S-1A

Quant Time: Jun 25 01:25:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

