



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062425\
Data File : VY022784.D
Acq On : 24 Jun 2025 08:49
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Title : SW846 8260
Last Update : Tue Jun 24 08:29:52 2025



AutoFind: Scans 1751, 1752, 1753; Background Corrected with Scan 1743

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	22.9	35827	PASS
75	95	30	60	55.8	87387	PASS
95	95	100	100	100.0	156613	PASS
96	95	5	9	6.8	10696	PASS
173	174	0.00	2	1.2	1461	PASS
174	95	50	100	79.8	125053	PASS
175	174	5	9	7.7	9683	PASS
176	174	95	101	96.9	121219	PASS
177	176	5	9	6.6	7948	PASS