

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
 Data File : VY018695.D
 Acq On : 25 Jun 2024 18:01
 Operator : SY/MD
 Sample : P3045-09
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :

Quant Time: Jun 26 02:40:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.777	168	98150	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	177973	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	162835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	70414	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	45959	46.509	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	93.020%		
35) Dibromofluoromethane	7.704	113	54403	50.110	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	100.220%		
50) Toluene-d8	10.173	98	208015	49.823	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	99.640%		
62) 4-Bromofluorobenzene	12.477	95	64511	48.932	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	97.860%		
Target Compounds						Qvalue
11) Tert butyl alcohol	4.942	59	342	1.924	ug/l #	88
16) Acetone	3.942	43	7867	39.131	ug/l #	83
17) Carbon Disulfide	4.180	76	6768	4.506	ug/l #	95
20) Methylene Chloride	4.698	84	8297	Below Cal		90
23) Vinyl Acetate	6.265	43	26	23.705	ug/l #	67
25) 2-Butanone	6.997	43	1072	3.557	ug/l #	69
31) Cyclohexane	7.765	56	5463	3.609	ug/l #	74
36) 1,1-Dichloropropene	7.777	75	7435	4.975	ug/l #	48
37) Ethyl Acetate	6.997	43	1072	1.522	ug/l #	67
38) Carbon Tetrachloride	7.771	117	9514	6.068	ug/l #	20
39) Methylcyclohexane	9.179	83	9323	5.092	ug/l	96
40) Benzene	8.149	78	268677	58.135	ug/l	100
42) 1,2-Dichloroethane	8.155	62	2066	1.713	ug/l #	71
49) 1,4-Dioxane	9.161	88	44	5.102	ug/l #	20
51) 4-Methyl-2-Pentanone	10.063	43	893	1.289	ug/l	94
52) Toluene	10.240	92	307811	103.736	ug/l	98
53) t-1,3-Dichloropropene	10.234	75	3597	2.630	ug/l #	1
55) 1,1,2-Trichloroethane	10.612	97	1095	1.256	ug/l #	16
59) 2-Hexanone	10.795	43	1533	3.160	ug/l #	22
67) Ethyl Benzene	11.587	91	107041	18.339	ug/l	92
68) m/p-Xylenes	11.697	106	182475	80.345	ug/l	98
69) o-Xylene	12.026	106	64857	29.696	ug/l	99
70) Styrene	12.038	104	7241	2.007	ug/l #	1
73) Isopropylbenzene	12.325	105	6761	1.315	ug/l	100
74) N-amyl acetate	12.105	43	2030	1.928	ug/l #	43
76) 1,2,3-Trichloropropane	12.477	75	33743	51.086	ug/l #	100
78) n-propylbenzene	12.666	91	14510	2.411	ug/l	98
80) 1,3,5-Trimethylbenzene	12.806	105	20117	4.937	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.477	75	33743	133.813	ug/l #	16
84) 1,2,4-Trimethylbenzene	13.117	105	50080	12.295	ug/l	99
95) Naphthalene	15.233	128	229114	107.408	ug/l	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
Data File : VY018695.D
Acq On : 25 Jun 2024 18:01
Operator : SY/MD
Sample : P3045-09
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :


Quant Time: Jun 26 02:40:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration








#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.777 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:168 Resp: 98150
Ion Ratio Lower Upper
168 100
99 54.0 39.1 58.7



#11
Tert butyl alcohol
Concen: 1.924 ug/l
RT: 4.942 min Scan# 512
Delta R.T. 0.006 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 59 Resp: 342
Ion Ratio Lower Upper
59 100
57 0.0 3.0 4.6#









#16
Acetone
Concen: 39.131 ug/l
RT: 3.942 min Scan# 346
Delta R.T. 0.012 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 7867
Ion Ratio Lower Upper
43 100
58 28.9 31.7 47.5#



#17
Carbon Disulfide
Concen: 4.506 ug/l
RT: 4.180 min Scan# 387
Delta R.T. 0.006 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 76 Resp: 6768
Ion Ratio Lower Upper
76 100
78 6.6 6.8 10.2#





#20

Methylene Chloride

Concen: Below Cal

RT: 4.698 min Scan# 41

Delta R.T. 0.000 min


Lab File: VY018695.D

Acq: 25 Jun 2024 18:01

Instrument :

MSVOA_Y

ClientSampleId :



Tgt Ion: 84 Resp: 8297



Ion Ratio Lower Upper

84 100

49 90.2 84.4 126.6

51 38.7 26.0 39.0

86 61.3 49.5 74.3



#23

Vinyl Acetate

Concen: 23.705 ug/l

RT: 6.265 min Scan# 729

Delta R.T. 0.220 min

Lab File: VY018695.D




Acq: 25 Jun 2024 18:01


Tgt Ion: 43 Resp: 26

Ion Ratio Lower Upper

43 100

86 0.0 10.6 16.0#









#25
2-Butanone
Concen: 3.557 ug/l
RT: 6.997 min Scan# 845
Delta R.T. 0.024 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 43 Resp: 1072
Ion Ratio Lower Upper
43 100
72 16.2 27.3 40.9#



#31
Cyclohexane
Concen: 3.609 ug/l
RT: 7.765 min Scan# 975
Delta R.T. -0.006 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 56 Resp: 5463
Ion Ratio Lower Upper
56 100
69 69.1 28.7 43.1#
84 111.8 77.7 116.5





#33

1,2-Dichloroethane-d4

Concen: 46.509 ug/l

RT: 8.130 min Scan# 1035

Delta R.T. 0.000 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 65 Resp: 45959

Ion Ratio Lower Upper

65 100

67 53.1 0.0 109.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.679 min Scan# 1125

Delta R.T. 0.000 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01




Tgt Ion: 114 Resp: 177973


Ion Ratio Lower Upper

114 100

63 17.0 0.0 35.0


88 14.1 0.0 27.2








#35
Dibromofluoromethane
Concen: 50.110 ug/l
RT: 7.704 min Scan# 910
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion:113 Resp: 54403
Ion Ratio Lower Upper
113 100
111 103.3 82.2 123.4
192 19.4 15.9 23.9



#36
1,1-Dichloropropene
Concen: 4.975 ug/l
RT: 7.777 min Scan# 977
Delta R.T. -0.128 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01


Tgt Ion: 75 Resp: 7435
Ion Ratio Lower Upper
75 100
110 9.3 19.0 57.0#
77 0.0 25.3 37.9#








#37
Ethyl Acetate
Concen: 1.522 ug/l
RT: 6.997 min Scan# 849
Delta R.T. -0.067 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 43 Resp: 1072
Ion Ratio Lower Upper
43 100
61 0.0 12.6 19.0#
70 0.0 7.8 11.8#



#38
Carbon Tetrachloride
Concen: 6.068 ug/l
RT: 7.771 min Scan# 976
Delta R.T. -0.116 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 117 Resp: 9514
Ion Ratio Lower Upper
117 100
119 9.3 75.8 113.6#
121 0.0 24.0 36.0#





#39

Methylcyclohexane

Concen: 5.092 ug/l

RT: 9.179 min Scan# 1

Delta R.T. 0.006 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01

Instrument :

MSVOA_Y

ClientSampleId :





Tgt Ion: 83 Resp: 9323

Ion Ratio Lower Upper

83 100

55 70.0 51.5 77.3

98 50.5 40.5 60.7



#40

Benzene

Concen: 58.135 ug/l

RT: 8.149 min Scan# 1038

Delta R.T. 0.000 min

Lab File: VY018695.D




Acq: 25 Jun 2024 18:01


Tgt Ion: 78 Resp: 268677

Ion Ratio Lower Upper

78 100

77 24.3 19.5 29.3









#42
1,2-Dichloroethane
Concen: 1.713 ug/l
RT: 8.155 min Scan# 1051
Delta R.T. -0.073 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 62 Resp: 2066
Ion Ratio Lower Upper
62 100
98 0.0 0.0 22.0



#49
1,4-Dioxane
Concen: 5.102 ug/l
RT: 9.161 min Scan# 1204
Delta R.T. -0.128 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 88 Resp: 44
Ion Ratio Lower Upper
88 100
43 1513.6 0.0 0.0#
58 0.0 49.3 73.9#









#50
Toluene-d8
Concen: 49.823 ug/l
RT: 10.173 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 98 Resp: 208015
Ion Ratio Lower Upper
98 100
100 64.3 52.0 78.0



#51
4-Methyl-2-Pentanone
Concen: 1.289 ug/l
RT: 10.063 min Scan# 1352
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 43 Resp: 893
Ion Ratio Lower Upper
43 100
58 40.9 35.7 53.5





#52

Toluene

Concen: 103.736 ug/l

RT: 10.240 min Scan# 11

Delta R.T. 0.000 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion: 92 Resp: 307811

Ion Ratio Lower Upper

92 100

91 170.7 139.0 208.6



#53

t-1,3-Dichloropropene

Concen: 2.630 ug/l

RT: 10.234 min Scan# 1380

Delta R.T. -0.225 min

Lab File: VY018695.D




Acq: 25 Jun 2024 18:01


Tgt Ion: 75 Resp: 3597

Ion Ratio Lower Upper

75 100


77 179.2 24.1 36.1#








#55
1,1,2-Trichloroethane
Concen: 1.256 ug/l
RT: 10.612 min Scan# 1442
Delta R.T. -0.024 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Instrument :
MSVOA_Y
ClientSampleId :







Tgt Ion: 97	Resp:	1095
Ion	Ratio	Lower Upper
97	100	
83	0.0	67.5 101.3#
85	0.0	42.9 64.3#
99	0.0	46.6 70.0#



#59
2-Hexanone
Concen: 3.160 ug/l
RT: 10.795 min Scan# 1472
Delta R.T. -0.030 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 43	Resp:	1533
Ion	Ratio	Lower Upper
43	100	
58	2.3	31.4 94.0#









#62
4-Bromofluorobenzene
Concen: 48.932 ug/l
RT: 12.477 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 95 Resp: 64511
Ion Ratio Lower Upper
95 100
174 89.3 0.0 175.6
176 86.6 0.0 171.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. 0.006 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion: 117 Resp: 162835
Ion Ratio Lower Upper
117 100
82 49.7 42.4 63.6
119 30.8 25.9 38.9









#67
Ethyl Benzene
Concen: 18.339 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 107041
Ion Ratio Lower Upper
91 100
106 34.7 24.5 36.7



#68
m/p-Xylenes
Concen: 80.345 ug/l
RT: 11.697 min Scan# 1620
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion:106 Resp: 182475
Ion Ratio Lower Upper
106 100
91 192.5 156.0 234.0





#69

o-Xylene

Concen: 29.696 ug/l

RT: 12.026 min Scan# 1

Delta R.T. 0.000 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01

Instrument :

MSVOA_Y





ClientSampleId :

Tgt Ion:106 Resp: 64857

Ion Ratio Lower Upper

106 100

91 206.4 103.6 310.8



#70

Styrene

Concen: 2.007 ug/l

RT: 12.038 min Scan# 1676

Delta R.T. 0.000 min

Lab File: VY018695.D

Acq: 25 Jun 2024 18:01




Tgt Ion:104 Resp: 7241


Ion Ratio Lower Upper

104 100

78 149.8 37.0 55.4#

103 152.4 44.2 66.2#









#72
 1,4-Dichlorobenzene-d4
 Concen: 50.000 ug/l
 RT: 13.422 min Scan# 1903
 Delta R.T. 0.000 min
 Lab File: VY018695.D
 Acq: 25 Jun 2024 18:01




Instrument :
 MSVOA_Y
 ClientSampleId :


Tgt Ion:152 Resp: 70414
 Ion Ratio Lower Upper
 152 100
 115 58.3 28.2 84.7
 150 161.2 0.0 345.6



#73
 Isopropylbenzene
 Concen: 1.315 ug/l
 RT: 12.325 min Scan# 1723
 Delta R.T. 0.000 min
 Lab File: VY018695.D
 Acq: 25 Jun 2024 18:01

Tgt Ion:105 Resp: 6761
 Ion Ratio Lower Upper
 105 100
 120 26.5 13.4 40.1









#74
N-amy1 acetate
Concen: 1.928 ug/l
RT: 12.105 min Scan# 1
Delta R.T. -0.036 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion:	43	Resp:	2030
Ion	Ratio	Lower	Upper
43	100		
70	4.9	41.4	62.2#
55	6.8	23.3	34.9#
61	0.0	22.6	33.8#



#76
1,2,3-Trichloropropane
Concen: 51.086 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.152 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Tgt Ion:	75	Resp:	33743
Ion	Ratio	Lower	Upper
75	100		
77	1.9	0.0	0.0#









#78
n-propylbenzene
Concen: 2.411 ug/l
RT: 12.666 min Scan# 1779
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Instrument :
MSVOA_Y
ClientSampleId :


Tgt Ion: 91 Resp: 14510
Ion Ratio Lower Upper
91 100
120 24.0 11.5 34.4



#80
1,3,5-Trimethylbenzene
Concen: 4.937 ug/l
RT: 12.806 min Scan# 1802
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01


Tgt Ion: 105 Resp: 20117
Ion Ratio Lower Upper
105 100
120 53.1 25.1 75.2








#81
trans-1,4-Dichloro-2-butene
Concen: 133.813 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. 0.104 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01


Instrument :
MSVOA_Y
ClientSampleId :





Tgt Ion: 75 Resp: 33743
Ion Ratio Lower Upper
75 100
53 0.0 66.8 100.2#
89 0.1 35.8 53.6#




#84
1,2,4-Trimethylbenzene
Concen: 12.295 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01




Tgt Ion: 105 Resp: 50080
Ion Ratio Lower Upper
105 100
120 45.4 23.2 69.6





#95
Naphthalene
Concen: 107.408 ug/l
RT: 15.233 min Scan# 21
Delta R.T. 0.000 min
Lab File: VY018695.D
Acq: 25 Jun 2024 18:01

Instrument :
MSVOA_Y
ClientSampleId :



Tgt Ion:128 Resp: 229114
Ion Ratio Lower Upper
128 100
127 13.0 10.6 16.0
129 10.6 8.6 13.0

