

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
 Data File : VY018686.D
 Acq On : 25 Jun 2024 14:17
 Operator : SY/MD
 Sample : P3021-01RE
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ARS20-008RE

Quant Time: Jun 26 02:36:41 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	66177	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.679	114	105095	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	84223	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	27972	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	24802	37.225	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	74.460%
35) Dibromofluoromethane	7.710	113	31428	49.022	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.040%
50) Toluene-d8	10.173	98	117870	47.809	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.477	95	28406	36.487	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	72.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
Data File : VY018686.D
Acq On : 25 Jun 2024 14:17
Operator : SY/MD
Sample : P3021-01RE
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ARS20-008RE

Quant Time: Jun 26 02:36:41 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

