

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
 Data File : VY018688.D
 Acq On : 25 Jun 2024 15:06
 Operator : SY/MD
 Sample : P3021-07RE
 Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ARS20-006RE

Quant Time: Jun 26 02:37:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 14 07:08:24 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.777	168	1306	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	2457	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	1487	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	129	50.000	ug/l	# 0.01

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	952	72.403	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	144.800%
35) Dibromofluoromethane	7.704	113	800	53.376	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.760%
50) Toluene-d8	10.173	98	2210	38.342	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	76.680%
62) 4-Bromofluorobenzene	12.471	95	144	7.912	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	15.820%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
Data File : VY018688.D
Acq On : 25 Jun 2024 15:06
Operator : SY/MD
Sample : P3021-07RE
Misc : 5.10g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ARS20-006RE

Quant Time: Jun 26 02:37:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Quant Title : SW846 8260
QLast Update : Fri Jun 14 07:08:24 2024
Response via : Initial Calibration

