



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062524\
Data File : VY018681.D
Acq On : 25 Jun 2024 09:18
Operator : SY/MD
Sample : BFB
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y061324S.M
Title : SW846 8260
Last Update : Fri Jun 14 07:08:24 2024



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1739

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.8	15082	PASS
75	95	30	60	51.9	43907	PASS
95	95	100	100	100.0	84589	PASS
96	95	5	9	6.6	5586	PASS
173	174	0.00	2	1.0	801	PASS
174	95	50	100	90.2	76328	PASS
175	174	5	9	7.5	5748	PASS
176	174	95	101	98.3	75008	PASS
177	176	5	9	6.4	4833	PASS