

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062525\
 Data File : VY022834.D
 Acq On : 25 Jun 2025 18:50
 Operator : SY/MD
 Sample : Q2423-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RT915

Quant Time: Jun 26 01:30:09 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	235402	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	449512	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	417191	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.340	152	160160	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	151081	57.504	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	115.000%
35) Dibromofluoromethane	7.628	113	146238	53.494	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.980%
50) Toluene-d8	10.103	98	538669	49.657	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.320%
62) 4-Bromofluorobenzene	12.401	95	172589	49.492	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	98.980%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.604	84	28453	9.073	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062525\
Data File : VY022834.D
Acq On : 25 Jun 2025 18:50
Operator : SY/MD
Sample : Q2423-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RT915

Quant Time: Jun 26 01:30:09 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

