

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062625\
 Data File : VY022842.D
 Acq On : 26 Jun 2025 12:10
 Operator : SY/MD
 Sample : Q2419-01
 Misc : 5.11g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-3-6-25-2025

Quant Time: Jun 27 01:26:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	214661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	369944	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	346828	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	170483	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	106198	44.326	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	88.660%		
35) Dibromofluoromethane	7.634	113	112990	50.222	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	100.440%		
50) Toluene-d8	10.103	98	444131	49.748	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	99.500%		
62) 4-Bromofluorobenzene	12.401	95	160445	55.905	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	111.820%		
Target Compounds						Qvalue
5) Bromomethane	2.610	94	3816	1.108	ug/l	94
8) Diethyl Ether	3.452	74	1695	1.415	ug/l	98
16) Acetone	3.866	43	1319	2.913	ug/l #	74
20) Methylene Chloride	4.616	84	16491	5.767	ug/l	89
31) Cyclohexane	7.713	56	5405	1.313	ug/l #	1
36) 1,1-Dichloropropene	7.707	75	18659	5.472	ug/l #	50
38) Carbon Tetrachloride	7.713	117	21687	6.027	ug/l #	19
39) Methylcyclohexane	9.109	83	8145	1.846	ug/l	95
55) 1,1,2-Trichloroethane	10.548	97	11394	6.317	ug/l #	21
56) Ethyl methacrylate	10.444	69	4243	1.748	ug/l #	1
58) 2-Chloroethyl Vinyl ether	9.731	63	80	15.714	ug/l #	48
59) 2-Hexanone	10.731	43	22728	21.509	ug/l #	30
69) o-Xylene	11.950	106	99470	20.395	ug/l	97
73) Isopropylbenzene	12.255	105	26582	2.108	ug/l	100
74) N-ethyl acetate	12.048	43	120936	42.732	ug/l #	60
75) 1,1,2,2-Tetrachloroethane	12.450	83	15059	7.411	ug/l #	25
76) 1,2,3-Trichloropropane	12.401	75	88410	50.796	ug/l #	100
78) n-propylbenzene	12.590	91	52886	3.474	ug/l	100
79) 2-Chlorotoluene	12.688	91	20070	2.334	ug/l #	46
80) 1,3,5-Trimethylbenzene	12.730	105	559753	54.995	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.401	75	88410	128.331	ug/l #	10
84) 1,2,4-Trimethylbenzene	13.041	105	18501	1.816	ug/l	99
85) sec-Butylbenzene	13.176	105	172057	12.740	ug/l	87
86) p-Isopropyltoluene	13.291	119	144343	12.843	ug/l	99
89) n-Butylbenzene	13.614	91	88548	8.376	ug/l #	55
90) Hexachloroethane	13.889	117	64211	28.685	ug/l #	16
92) 1,2-Dibromo-3-Chloropr...	14.297	75	2369	6.875	ug/l #	27
93) 1,2,4-Trichlorobenzene	14.919	180	4795	1.679	ug/l #	1
95) Naphthalene	15.139	128	29739	5.759	ug/l #	69
96) 1,2,3-Trichlorobenzene	15.340	180	6096	2.470	ug/l #	1





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion:168 Resp: 214661
Ion Ratio Lower Upper
168 100
99 56.8 44.3 66.5



#5
Bromomethane
Concen: 1.108 ug/l
RT: 2.610 min Scan# 146
Delta R.T. 0.012 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 94 Resp: 3816
Ion Ratio Lower Upper
94 100
96 98.2 74.0 111.0









#8
Diethyl Ether
Concen: 1.415 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.013 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 74 Resp: 1695
Ion Ratio Lower Upper
74 100
45 93.3 45.9 137.6



#16
Acetone
Concen: 2.913 ug/l
RT: 3.866 min Scan# 352
Delta R.T. -0.013 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 43 Resp: 1319
Ion Ratio Lower Upper
43 100
58 16.0 24.0 36.0#









#20
Methylene Chloride
Concen: 5.767 ug/l
RT: 4.616 min Scan# 41
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025

Tgt Ion: 84 Resp: 16491




Ion	Ratio	Lower	Upper
84	100		
49	113.2	101.9	152.9
51	35.9	29.8	44.8
86	52.8	51.3	76.9




#31
Cyclohexane
Concen: 1.313 ug/l
RT: 7.713 min Scan# 983
Delta R.T. 0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 56 Resp: 5405

Ion	Ratio	Lower	Upper
56	100		
69	160.9	27.0	40.6#
84	143.3	72.6	109.0#





#33

1,2-Dichloroethane-d4

Concen: 44.326 ug/l

RT: 8.061 min Scan# 110

Delta R.T. -0.006 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10

Instrument :

MSVOA_Y

ClientSampleId :





EO-3-6-25-2025

Tgt Ion: 65 Resp: 106198

Ion Ratio Lower Upper

65 100

67 52.3 0.0 103.4



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.615 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10




Tgt Ion:114 Resp: 369944


Ion Ratio Lower Upper

114 100

63 20.1 0.0 40.8

88 16.1 0.0 27.8









#35
Dibromofluoromethane
Concen: 50.222 ug/l
RT: 7.634 min Scan# 91
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion:113 Resp: 112990
Ion Ratio Lower Upper
113 100
111 103.3 81.1 121.7
192 19.1 14.2 21.2



#36
1,1-Dichloropropene
Concen: 5.472 ug/l
RT: 7.707 min Scan# 982
Delta R.T. -0.134 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 75 Resp: 18659
Ion Ratio Lower Upper
75 100
110 8.7 16.9 50.7#
77 0.0 24.8 37.2#









#38
Carbon Tetrachloride
Concen: 6.027 ug/l
RT: 7.713 min Scan# 91
Delta R.T. -0.110 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion:117 Resp: 21687
Ion Ratio Lower Upper
117 100
119 8.4 75.0 112.6#
121 0.0 25.0 37.4#



#39
Methylcyclohexane
Concen: 1.846 ug/l
RT: 9.109 min Scan# 1212
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 83 Resp: 8145
Ion Ratio Lower Upper
83 100
55 70.2 58.1 87.1
98 43.9 39.0 58.6









#50
Toluene-d8
Concen: 49.748 ug/l
RT: 10.103 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 98 Resp: 444131
Ion Ratio Lower Upper
98 100
100 64.0 51.4 77.0



#55
1,1,2-Trichloroethane
Concen: 6.317 ug/l
RT: 10.548 min Scan# 1448
Delta R.T. -0.025 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 97 Resp: 11394
Ion Ratio Lower Upper
97 100
83 10.1 68.1 102.1#
85 4.4 48.2 72.2#
99 0.0 49.5 74.3#









#56
Ethyl methacrylate
Concen: 1.748 ug/l
RT: 10.444 min Scan# 1429
Delta R.T. -0.000 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 69 Resp: 4243
Ion Ratio Lower Upper
69 100
41 185.1 52.2 78.2#
39 105.8 36.7 55.1#



#58
2-Chloroethyl Vinyl ether
Concen: 15.714 ug/l
RT: 9.731 min Scan# 1314
Delta R.T. 0.018 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 63 Resp: 80
Ion Ratio Lower Upper
63 100
106 0.0 21.6 32.4#









#59
2-Hexanone
Concen: 21.509 ug/l
RT: 10.731 min Scan# 1482
Delta R.T. -0.031 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 43 Resp: 22728
Ion Ratio Lower Upper
43 100
58 3.8 27.3 81.8#



#62
4-Bromofluorobenzene
Concen: 55.905 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 95 Resp: 160445
Ion Ratio Lower Upper
95 100
174 81.9 0.0 170.0
176 80.0 0.0 166.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 1590
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion:117 Resp: 346828
Ion Ratio Lower Upper
117 100
82 57.0 44.6 66.8
119 32.2 25.4 38.0



#69
o-Xylene
Concen: 20.395 ug/l
RT: 11.950 min Scan# 1678
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion:106 Resp: 99470
Ion Ratio Lower Upper
106 100
91 216.4 105.8 317.3





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1906

Delta R.T. -0.000 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10

Instrument :

MSVOA_Y

ClientSampleId :

EO-3-6-25-2025





Tgt Ion:152 Resp: 170483

Ion Ratio Lower Upper

152 100

115 59.6 28.9 86.7

150 153.1 0.0 349.6



#73

Isopropylbenzene

Concen: 2.108 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. -0.000 min

Lab File: VY022842.D




Acq: 26 Jun 2025 12:10


Tgt Ion:105 Resp: 26582

Ion Ratio Lower Upper

105 100

120 25.9 13.0 38.9









#74
N-amy1 acetate
Concen: 42.732 ug/l
RT: 12.048 min Scan# 1
Delta R.T. -0.025 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 43 Resp: 120936
Ion Ratio Lower Upper
43 100
70 11.3 34.2 51.2#
55 16.4 18.2 27.4#
61 0.0 19.0 28.6#



#75
1,1,2,2-Tetrachloroethane
Concen: 7.411 ug/l
RT: 12.450 min Scan# 1760
Delta R.T. -0.055 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 83 Resp: 15059
Ion Ratio Lower Upper
83 100
131 0.0 4.9 14.7#
85 0.0 32.6 97.8#









#76
1,2,3-Trichloropropane
Concen: 50.796 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. -0.153 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 75 Resp: 88410
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0



#78
n-propylbenzene
Concen: 3.474 ug/l
RT: 12.590 min Scan# 1783
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 91 Resp: 52886
Ion Ratio Lower Upper
91 100
120 21.5 10.8 32.4





#79

2-Chlorotoluene

Concen: 2.334 ug/l

RT: 12.688 min Scan# 1799

Delta R.T. 0.006 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10

Instrument :

MSVOA_Y

ClientSampleId :





EO-3-6-25-2025

Tgt Ion: 91 Resp: 20070

Ion Ratio Lower Upper

91 100

126 2.9 17.0 51.0#



#80

1,3,5-Trimethylbenzene

Concen: 54.995 ug/l

RT: 12.730 min Scan# 1806

Delta R.T. -0.006 min

Lab File: VY022842.D




Acq: 26 Jun 2025 12:10


Tgt Ion: 105 Resp: 559753

Ion Ratio Lower Upper

105 100

120 49.0 24.1 72.4









#81
trans-1,4-Dichloro-2-butene
Concen: 128.331 ug/l
RT: 12.401 min Scan# 1752
Delta R.T. 0.097 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 75 Resp: 88410
Ion Ratio Lower Upper
75 100
53 0.0 78.5 117.7#
89 0.1 35.7 53.5#



#84
1,2,4-Trimethylbenzene
Concen: 1.816 ug/l
RT: 13.041 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 105 Resp: 18501
Ion Ratio Lower Upper
105 100
120 46.0 22.7 68.1









#85
 sec-Butylbenzene
 Concen: 12.740 ug/l
 RT: 13.176 min Scan# 1878
 Delta R.T. -0.000 min
 Lab File: VY022842.D
 Acq: 26 Jun 2025 12:10




Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-3-6-25-2025


Tgt Ion:105 Resp: 172057
 Ion Ratio Lower Upper
 105 100
 134 25.3 9.7 28.9



#86
 p-Isopropyltoluene
 Concen: 12.843 ug/l
 RT: 13.291 min Scan# 1898
 Delta R.T. -0.000 min
 Lab File: VY022842.D
 Acq: 26 Jun 2025 12:10

Tgt Ion:119 Resp: 144343
 Ion Ratio Lower Upper
 119 100
 134 26.2 13.1 39.1
 91 22.5 11.6 34.8









#89
n-Butylbenzene
Concen: 8.376 ug/l
RT: 13.614 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10




Instrument :
MSVOA_Y
ClientSampleId :
EO-3-6-25-2025


Tgt Ion: 91 Resp: 88548
Ion Ratio Lower Upper
91 100
92 82.0 25.9 77.8#
134 0.0 12.0 36.1#



#90
Hexachloroethane
Concen: 28.685 ug/l
RT: 13.889 min Scan# 1996
Delta R.T. 0.006 min
Lab File: VY022842.D
Acq: 26 Jun 2025 12:10

Tgt Ion: 117 Resp: 64211
Ion Ratio Lower Upper
117 100
201 0.3 34.0 102.0#









#92
 1,2-Dibromo-3-Chloropropane
 Concen: 6.875 ug/l
 RT: 14.297 min Scan# 2058
 Delta R.T. 0.024 min
 Lab File: VY022842.D
 Acq: 26 Jun 2025 12:10




Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-3-6-25-2025


Tgt Ion: 75 Resp: 2369
 Ion Ratio Lower Upper
 75 100
 155 52.3 39.9 119.6
 157 0.0 54.0 162.0#



#93
 1,2,4-Trichlorobenzene
 Concen: 1.679 ug/l
 RT: 14.919 min Scan# 2165
 Delta R.T. -0.000 min
 Lab File: VY022842.D
 Acq: 26 Jun 2025 12:10

Tgt Ion: 180 Resp: 4795
 Ion Ratio Lower Upper
 180 100
 182 123.6 48.4 145.3
 145 1903.0 15.5 46.5#





#95

Naphthalene

Concen: 5.759 ug/l

RT: 15.139 min Scan# 21

Delta R.T. -0.006 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10

Instrument :

MSVOA_Y

ClientSampleId :

EO-3-6-25-2025





Tgt Ion:128 Resp: 29739

Ion Ratio Lower Upper

128 100

127 0.0 10.4 15.6#

129 0.0 8.9 13.3#



#96

1,2,3-Trichlorobenzene

Concen: 2.470 ug/l

RT: 15.340 min Scan# 2234

Delta R.T. 0.012 min

Lab File: VY022842.D

Acq: 26 Jun 2025 12:10

Tgt Ion:180 Resp: 6096

Ion Ratio Lower Upper

180 100

182 76.1 47.1 141.3

145 799.5 15.2 45.6#

