

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062725\
 Data File : VY022859.D
 Acq On : 27 Jun 2025 14:09
 Operator : SY/MD
 Sample : Q2436-04
 Misc : 7.25g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-86

Quant Time: Jun 27 23:18:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	145217	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	257521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	236239	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	89826	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	82314	50.787	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.580%
35) Dibromofluoromethane	7.640	113	82993	52.993	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.980%
50) Toluene-d8	10.109	98	299626	48.213	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.420%
62) 4-Bromofluorobenzene	12.407	95	100986	50.549	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062725\
Data File : VY022859.D
Acq On : 27 Jun 2025 14:09
Operator : SY/MD
Sample : Q2436-04
Misc : 7.25g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-86

Quant Time: Jun 27 23:18:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

