

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062725\
 Data File : VY022864.D
 Acq On : 27 Jun 2025 16:06
 Operator : SY/MD
 Sample : Q2436-09
 Misc : 5.48g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-99

Quant Time: Jun 27 23:19:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	301121	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	558142	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	536522	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	230605	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	168750	50.211	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.420%
35) Dibromofluoromethane	7.634	113	172035	50.683	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.360%
50) Toluene-d8	10.109	98	673924	50.034	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.060%
62) 4-Bromofluorobenzene	12.402	95	234963	54.265	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	108.520%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY062725\
Data File : VY022864.D
Acq On : 27 Jun 2025 16:06
Operator : SY/MD
Sample : Q2436-09
Misc : 5.48g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-99

Quant Time: Jun 27 23:19:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

