

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063025\
 Data File : VY022882.D
 Acq On : 30 Jun 2025 16:02
 Operator : SY/MD
 Sample : Q2449-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RBR200032

Quant Time: Jul 01 03:39:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	298735	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	565547	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	561474	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	246923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	174934	52.467	ug/l	-0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.940%	
35) Dibromofluoromethane	7.628	113	176723	51.382	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.760%	
50) Toluene-d8	10.103	98	691265	50.650	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.300%	
62) 4-Bromofluorobenzene	12.402	95	249495	56.866	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 113.740%	
Target Compounds						Qvalue
20) Methylene Chloride	4.610	84	26858	6.749	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY063025\
Data File : VY022882.D
Acq On : 30 Jun 2025 16:02
Operator : SY/MD
Sample : Q2449-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RBR200032

Quant Time: Jul 01 03:39:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

