

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070125\
 Data File : VY022895.D
 Acq On : 01 Jul 2025 12:34
 Operator : SY/MD
 Sample : Q2457-01RE
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-2-062725RE

Quant Time: Jul 02 04:29:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	159831	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	281268	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	263568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	109145	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	90148	50.535	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.060%
35) Dibromofluoromethane	7.628	113	88860	51.949	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.900%
50) Toluene-d8	10.103	98	332025	48.916	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.840%
62) 4-Bromofluorobenzene	12.401	95	110833	50.794	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.580%
Target Compounds						
20) Methylene Chloride	4.604	84	8538	4.010	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070125\
Data File : VY022895.D
Acq On : 01 Jul 2025 12:34
Operator : SY/MD
Sample : Q2457-01RE
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OK-2-062725RE

Quant Time: Jul 02 04:29:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

