

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070125\
 Data File : VY022897.D
 Acq On : 01 Jul 2025 13:20
 Operator : SY/MD
 Sample : Q2458-03
 Misc : 5.92g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-68

Quant Time: Jul 02 02:46:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	309337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.610	114	580195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	586876	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	251388	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	176384	51.089	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.180%
35) Dibromofluoromethane	7.634	113	184079	52.170	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.340%
50) Toluene-d8	10.103	98	716167	51.149	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.300%
62) 4-Bromofluorobenzene	12.402	95	256602	57.010	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	114.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070125\
Data File : VY022897.D
Acq On : 01 Jul 2025 13:20
Operator : SY/MD
Sample : Q2458-03
Misc : 5.92g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-68

Quant Time: Jul 02 02:46:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

