

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
 Data File : VY022915.D
 Acq On : 02 Jul 2025 16:20
 Operator : SY/MD
 Sample : Q2458-08
 Misc : 4.91g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-63

Quant Time: Jul 03 02:24:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	278456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	537398	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	515031	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	195752	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	166821	53.677	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.360%
35) Dibromofluoromethane	7.634	113	166540	50.958	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.920%
50) Toluene-d8	10.103	98	648984	50.042	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.080%
62) 4-Bromofluorobenzene	12.401	95	212971	51.084	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	102.160%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
Data File : VY022915.D
Acq On : 02 Jul 2025 16:20
Operator : SY/MD
Sample : Q2458-08
Misc : 4.91g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-63

Quant Time: Jul 03 02:24:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

