

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
 Data File : VY022918.D
 Acq On : 02 Jul 2025 17:30
 Operator : SY/MD
 Sample : Q2473-02
 Misc : 8.47g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PIT#2

Quant Time: Jul 03 02:25:43 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	300149	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	579446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	556091	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	220138	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	181616	54.214	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 108.420%	
35) Dibromofluoromethane	7.634	113	125921	35.733	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 71.460%	
50) Toluene-d8	10.103	98	710362	50.800	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.600%	
62) 4-Bromofluorobenzene	12.402	95	233105	51.856	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 103.720%	
Target Compounds						
16) Acetone	3.873	43	3990	6.302	ug/l	Qvalue 92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070225\
Data File : VY022918.D
Acq On : 02 Jul 2025 17:30
Operator : SY/MD
Sample : Q2473-02
Misc : 8.47g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PIT#2

Quant Time: Jul 03 02:25:43 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

