

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
 Data File : VY022950.D
 Acq On : 07 Jul 2025 11:18
 Operator : SY/MD
 Sample : Q2507-03
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-04-7.3-2025-VOC

Quant Time: Jul 08 01:44:31 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
 Quant Title : SW846 8260
 QLast Update : Tue Jun 24 08:29:52 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	169827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	291091	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	256345	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	96874	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	83563	44.086	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	88.180%
35) Dibromofluoromethane	7.634	113	87585	49.475	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.960%
50) Toluene-d8	10.103	98	342304	48.729	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.460%
62) 4-Bromofluorobenzene	12.402	95	102452	45.369	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	90.740%
Target Compounds						
20) Methylene Chloride	4.616	84	19723	8.717	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070725\
Data File : VY022950.D
Acq On : 07 Jul 2025 11:18
Operator : SY/MD
Sample : Q2507-03
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-04-7.3-2025-VOC

Quant Time: Jul 08 01:44:31 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y062325S.M
Quant Title : SW846 8260
QLast Update : Tue Jun 24 08:29:52 2025
Response via : Initial Calibration

