

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
 Data File : VY009481.D
 Acq On : 08 Jul 2022 14:37
 Operator : KP/MD
 Sample : N3595-01
 Misc : 6.56g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20220487A-5A

Quant Time: Jul 11 02:14:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.782	168	199358	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	330900	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	297358	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	129485	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	70216	50.000	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.000%
35) Dibromofluoromethane	7.709	113	61439	35.288	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	70.580%
50) Toluene-d8	10.172	98	233977	48.823	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	97.640%
62) 4-Bromofluorobenzene	12.477	95	75914	30.888	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	61.780%
Target Compounds						Qvalue
16) Acetone	3.936	43	13110	27.033	ug/l	92
17) Carbon Disulfide	4.186	76	212267	40.125	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
Data File : VY009481.D
Acq On : 08 Jul 2022 14:37
Operator : KP/MD
Sample : N3595-01
Misc : 6.56g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20220487A-5A

Quant Time: Jul 11 02:14:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

