

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
 Data File : VY009494.D
 Acq On : 08 Jul 2022 19:43
 Operator : KP/MD
 Sample : N3595-14
 Misc : 6.64g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 20220491B-9B

Quant Time: Jul 11 02:16:30 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	210979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	317776	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.483	117	348213	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	156625	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	72912	48.920	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.840%
35) Dibromofluoromethane	7.710	113	64433	38.536	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	77.080%
50) Toluene-d8	10.173	98	242061	53.345	ug/l	0.00
Spiked Amount	50.000	Range	78 - 125	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.477	95	112642	47.725	ug/l	0.00
Spiked Amount	50.000	Range	50 - 146	Recovery	=	95.460%
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	9289	18.099	ug/l #	84
17) Carbon Disulfide	4.180	76	215712	38.706	ug/l	99
18) Methyl Acetate	4.466	43	1940	1.287	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
Data File : VY009494.D
Acq On : 08 Jul 2022 19:43
Operator : KP/MD
Sample : N3595-14
Misc : 6.64g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
20220491B-9B

Quant Time: Jul 11 02:16:30 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

