

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
 Data File : VY009477.D
 Acq On : 08 Jul 2022 10:28
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 11 02:13:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.383	0.381	0.5	101	0.00
3 P	Chloromethane	0.428	0.441	-3.0	98	0.00
4 C	Vinyl Chloride	0.488	0.501	-2.7#	95	0.00
5 T	Bromomethane	0.350	0.319	8.9	93	0.00
6 T	Chloroethane	0.285	0.270	5.3	96	0.00
7 T	Trichlorofluoromethane	0.651	0.652	-0.2	98	0.00
8 T	Diethyl Ether	0.249	0.198	20.5	77	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.441	0.436	1.1	98	0.00
10 T	Methyl Iodide	0.523	0.580	-10.9	100	0.00
11 T	Tert butyl alcohol	0.077	0.053	31.2#	87	0.00
12 CM	1,1-Dichloroethene	0.409	0.423	-3.4#	98	0.00
13 T	Acrolein	0.031	0.040	-29.0#	103	0.00
14 T	Allyl chloride	0.704	0.760	-8.0	97	0.00
15 T	Acrylonitrile	0.137	0.149	-8.8	104	0.00
16 T	Acetone	0.122	0.125	-2.5	77	0.00
17 T	Carbon Disulfide	1.151	1.319	-14.6	96	0.00
18 T	Methyl Acetate	0.357	0.364	-2.0	110	0.00
19 T	Methyl tert-butyl Ether	1.238	1.289	-4.1	102	0.00
20 T	Methylene Chloride	0.685	0.501	26.9#	89	0.00
21 T	trans-1,2-Dichloroethene	0.479	0.509	-6.3	99	0.00
22 T	Diisopropyl ether	1.496	1.658	-10.8	99	0.00
23 T	Vinyl Acetate	0.957	1.120	-17.0	99	0.00
24 P	1,1-Dichloroethane	0.844	0.888	-5.2	101	0.00
25 T	2-Butanone	0.197	0.208	-5.6	91	0.00
26 T	2,2-Dichloropropane	0.832	0.788	5.3	95	0.00
27 T	cis-1,2-Dichloroethene	0.564	0.584	-3.5	101	0.00
28 T	Bromochloromethane	0.355	0.378	-6.5	101	0.00
29 T	Tetrahydrofuran	0.127	0.137	-7.9	101	0.00
30 C	Chloroform	0.923	0.916	0.8#	101	0.00
31 T	Cyclohexane	0.820	0.843	-2.8	96	0.00
32 T	1,1,1-Trichloroethane	0.814	0.812	0.2	99	0.00
33 S	1,2-Dichloroethane-d4	0.427	0.324	24.1	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.263	0.229	12.9	101	0.00
36 T	1,1-Dichloropropene	0.428	0.447	-4.4	98	0.00
37 T	Ethyl Acetate	0.268	0.285	-6.3	101	0.00
38 T	Carbon Tetrachloride	0.458	0.461	-0.7	97	0.00
39 T	Methylcyclohexane	0.514	0.567	-10.3	97	0.00
40 TM	Benzene	1.258	1.307	-3.9	98	0.00
41 T	Methacrylonitrile	0.139	0.176	-26.6#	115	0.00
42 TM	1,2-Dichloroethane	0.378	0.378	0.0	98	0.00
43 T	Isopropyl Acetate	0.501	0.528	-5.4	100	0.00
44 TM	Trichloroethene	0.355	0.365	-2.8	98	0.00
45 C	1,2-Dichloropropane	0.318	0.332	-4.4#	99	0.00
46 T	Dibromomethane	0.187	0.197	-5.3	101	0.00
47 T	Bromodichloromethane	0.445	0.444	0.2	99	0.00
48 T	Methyl methacrylate	0.216	0.244	-13.0	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
 Data File : VY009477.D
 Acq On : 08 Jul 2022 10:28
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Jul 11 02:13:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 08 11:00:56 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	102	0.00
50 S	Toluene-d8	0.904	0.664	26.5#	98	0.00
51 T	4-Methyl-2-Pentanone	0.263	0.282	-7.2	100	0.00
52 CM	Toluene	0.801	0.842	-5.1#	96	0.00
53 T	t-1,3-Dichloropropene	0.469	0.485	-3.4	97	0.00
54 T	cis-1,3-Dichloropropene	0.522	0.537	-2.9	97	0.00
55 T	1,1,2-Trichloroethane	0.270	0.273	-1.1	96	0.00
56 T	Ethyl methacrylate	0.358	0.387	-8.1	96	0.00
57 T	1,3-Dichloropropane	0.444	0.460	-3.6	98	0.00
58 T	2-Chloroethyl Vinyl ether	0.175	0.191	-9.1	99	0.00
59 T	2-Hexanone	0.183	0.202	-10.4	96	0.00
60 T	Dibromochloromethane	0.326	0.346	-6.1	100	0.00
61 T	1,2-Dibromoethane	0.258	0.266	-3.1	97	0.00
62 S	4-Bromofluorobenzene	0.371	0.358	3.5	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.418	0.422	-1.0	98	0.00
65 PM	Chlorobenzene	0.982	0.983	-0.1	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.369	0.366	0.8	98	0.00
67 C	Ethyl Benzene	1.714	1.767	-3.1#	99	0.00
68 T	m/p-Xylenes	0.669	0.688	-2.8	98	0.00
69 T	o-Xylene	0.641	0.650	-1.4	96	0.00
70 T	Styrene	1.070	1.114	-4.1	99	0.00
71 P	Bromoform	0.235	0.249	-6.0	102	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.419	3.536	-3.4	99	0.00
74 T	N-amyl acetate	0.973	1.006	-3.4	103	0.00
75 P	1,1,2,2-Tetrachloroethane	0.685	0.715	-4.4	102	0.00
76 T	1,2,3-Trichloropropane	0.529	0.521	1.5	89	0.00
77 T	Bromobenzene	0.832	0.842	-1.2	99	0.00
78 T	n-propylbenzene	4.131	4.261	-3.1	99	0.00
79 T	2-Chlorotoluene	2.313	2.391	-3.4	99	0.00
80 T	1,3,5-Trimethylbenzene	2.858	2.979	-4.2	100	0.00
81 T	trans-1,4-Dichloro-2-butene	0.259	0.270	-4.2	99	0.00
82 T	4-Chlorotoluene	2.424	2.482	-2.4	100	0.00
83 T	tert-Butylbenzene	2.543	2.625	-3.2	100	0.00
84 T	1,2,4-Trimethylbenzene	2.845	2.911	-2.3	99	0.00
85 T	sec-Butylbenzene	3.745	3.897	-4.1	101	0.00
86 T	p-Isopropyltoluene	3.122	3.220	-3.1	100	0.00
87 T	1,3-Dichlorobenzene	1.643	1.654	-0.7	100	0.00
88 T	1,4-Dichlorobenzene	1.663	1.665	-0.1	101	0.00
89 T	n-Butylbenzene	2.915	3.041	-4.3	102	0.00
90 T	Hexachloroethane	0.613	0.619	-1.0	100	0.00
91 T	1,2-Dichlorobenzene	1.506	1.520	-0.9	102	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.130	-8.3	107	0.00
93 T	1,2,4-Trichlorobenzene	0.918	0.928	-1.1	102	0.00
94 T	Hexachlorobutadiene	0.505	0.509	-0.8	103	0.00
95 T	Naphthalene	1.899	2.070	-9.0	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY070822\
Data File : VY009477.D
Acq On : 08 Jul 2022 10:28
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Jul 11 02:13:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y070722S.M
Quant Title : SW846 8260
QLast Update : Fri Jul 08 11:00:56 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.795	0.816	-2.6	103	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6